

pH LEVELS, CONDUCTIVITY & DENSITY

Natrix

HR2-116

Reagent Number	Measured pH Levels	Conductivity (mS or μ S)	Density (g/cm ³)	Reagent Number	Measured pH Levels	Conductivity (mS or μ S)	Density (g/cm ³)
1	5.6	78.9 mS	1.168	25	6.8	11.5 mS	1.057
2	6.0	312.1 mS	1.163	26	6.5	25.1 mS	1.034
3	5.9	4.9 mS	1.007	27	6.9	7.5 mS	1.055
4	5.3	18.4 mS	1.028	28	7.0	78.2 mS	1.142
5	5.3	23.0 mS	1.018	29	7.0	159.1 mS	1.091
6	5.3	10.3 mS	1.041	30	6.9	177.0 mS	1.111
7	5.8	3.3 mS	0.982	31	6.7	7.9 mS	1.042
8	6.0	60.3 mS	1.026	32	6.7	16.1 mS	1.010
9	5.8	11.1 mS	1.023	33	6.6	11.9 mS	1.003
10	5.8	1761.7 μ S	1.008	34	6.7	10.7 mS	1.006
11	5.6	69.9 mS	1.090	35	6.7	15.7 mS	1.016
12	5.5	80.9 mS	1.154	36	6.7	13.5 mS	1.025
13	5.9	183.6 mS	1.118	37	6.8	15.6 mS	1.046
14	6.1	12.7 mS	0.988	38	6.9	25.3 mS	1.025
15	5.9	9.5 mS	1.003	39	6.8	14.0 mS	1.013
16	6.4	3.0 mS	1.003	40	7.3	180.0 mS	1.109
17	6.0	22.4 mS	1.028	41	7.5	13.8 mS	1.020
18	5.9	75.4 mS	1.114	42	7.5	5.5 mS	0.992
19	6.2	237.7 mS	1.136	43	7.5	7.6 mS	1.000
20	6.5	10.0 mS	0.991	44	7.6	26.7 mS	1.028
21	6.4	20.5 mS	1.010	45	7.9	192.1 mS	1.122
22	6.5	6.8 mS	1.034	46	8.3	1879.3 μ S	0.996
23	6.4	21.5 mS	1.027	47	8.4	6.4 mS	1.057
24	6.5	15.8 mS	1.022	48	8.1	11.4 mS	1.053

Copyright 2006 Hampton Research Corp.